



NCP - A

2161

Memorandum of Understanding

The memorandum of understanding is prepared for the transfer of HPTLC Applicator model AS30 (DESAGA SARSTEDT-GRUPPE GMBH) from Labchrom Scientific LLP to Nirmala College of Pharmacy.

Signed by Parties:-

- (A) Mr. Rajendra Barabde (Partner) - M/s. Labchrom Scientific LLP (consignor) , Unit 253, Udyog Bhavan , Sonawala Road, Goregaon East, Mumbai – 400063, Maharashtra
(B) **Prof. Dr. Badmanaban. R, M. Pharm., PhD (Principal)** (consignee)
Nirmala College of Pharmacy, Muvattupuzha P.O, Ernakulam District, Kerala-686661


Purpose & Responsibilities: -

Shifting of AS30 Applicator from Labchrom Mumbai office to Nirmala College of Pharmacy, Kerala for period of 2 years (till 30th June 2025). The ownership of AS30 applicator will remain with Labchrom Scientific LLP only.

It is mutually decided that HPTLC applicator model AS30 owned by M/s. Labchrom Scientific LLP, will ship to Nirmala College of Pharmacy Muvattupuzha P.O, Ernakulam District, Kerala-686661 for training purpose.

- Transportation and installation expenses will be borne by Labchrom Scientific LLP
- Nitrogen Gas cylinder and normal glass type chamber, HPTLC plates, UV box etc.. arrangements will be done by Nirmala College of Pharmacy.
- Labchrom will not charge any kind of installation and service charges during this period and the training to students at Nirmala College of Pharmacy will be given free of charge. In exchange **Labchrom will use expertise of Prof. Dr. Badmanaban. R,** for the purpose of free demonstration to their clients, if any.
- During these two years if any trouble or complaint occurred in Applicator then Labchrom will take care of the same - Train Ticket 3AC, Accommodation, if necessary, can be borne by Nirmala College of Pharmacy for the engineer who visits.

For LABCHROM SCIENTIFIC LLP



Rajendra Barabde
Partner
Date: -



For NIRMALA COLLEGE OF PHARMACY


18.07.23

Prof. Dr. Badmanaban. R
Principal
Date:- 18.07.23

Dr. BADMANABAN. R
PRINCIPAL
Nirmala College of Pharmacy
Muvattupuzha, Ernakulam (Dist.)
Kerala - 686 661



HPTLC Applicator Model AS30 box contains:-

- 1) HPTLC Applicator Model AS30 Qty 1.
- 2) AS30 Tower Qty. 1
- 3) Transformer Qty. 1 (220 to 110V)
- 4) Filling Syringe Qty. 2
- 5) Dosing Syringe Qty. 1
- 6) Power Cable Qty. 1
- 7) HPTLC Plates Qty. 20 (1 box)

TECHNICAL SERVICE VISIT REPORT

MASTER COPY

Labchrom Scientific LLP

253, 2nd Floor, Udyog Bhavan, Sonawala Road, Goregaon East, Mumbai - 400063 Maharashtra (India)
Tel.: +91 22 26851181 / 2 / 3 Fax: +91 22 26851183 , Email : mail@labchrom.com , website : www.labchrom.com

Name of Customer & Address with Pin : **NIRMALA COLLEGE OF PHARMACY, Muvattupuzha P.O., Ernakulam District, Kerala-686661, India.**

Contact person Name & Designation / Dept:- **Dr. Badmanaban / Principal**
Mobile No :- _____ Email id :- _____

Date	Instrument / System Configuration	Model	Serial No.	
17/07/23	Applicator	AS30	04626	Warranty <input type="checkbox"/>
18/07/23				AMC <input type="checkbox"/>
				Courtesy <input checked="" type="checkbox"/>
				Chargeable <input type="checkbox"/>

Nature of Problem :- Installation and demonstration of Desaga AS30 instruments .


Observation :-

Action taken :-

1. Unpacked the Package, checked the physical condition ok.
2. Installed the HPTLC AS 30 instruments and software successfully.
3. Checked the operation of HPTLC AS 30 instruments
4. Demonstrated the instruments operation and software operation.
5. Injected the Sample found instruments working satisfactory.
6. Installation and demonstration work completed successfully.

Parts Recommended :-

Customer Remarks :-

Service Engineer Name & Signature	Customer Name, Signature, Stamp & Date
S. Annamalai S.A.	 18.07.23 Dr. BADMANABAN. R PRINCIPAL Nirmala College of Pharmacy Muvattupuzha, Ernakulam (Dist.) Kerala - 686 661